



Accredited Laboratory

A2LA has accredited

GREENVILLE SCALE CO., INC.

Taylors, SC

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 26th day of January 2023.

A blue ink signature of Mr. Trace McInturff.

Mr. Trace McInturff, Vice President, Accreditation Services
For the Accreditation Council
Certificate Number 1524.01
Valid to January 31, 2025

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017
& ANSI/NCSL Z540-1-1994

GREENVILLE SCALE CO., INC.
149 Landmark Drive
Taylors, SC 29687
George Williamson Phone: 864 244 4723

CALIBRATION

Valid To: January 31, 2025

Certificate Number: 1524.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Mechanical

Parameter/Equipment	Range	CMC ^{2,4} (±)	Comments
Scales & Balances ³ –	Up to 1 g	0.04 mg + 0.6R	Handbook 44 with: ASTM Class 1 weights
	(1 to 10) g	0.06 mg + 0.6R	
	(10 to 30) g	0.09 mg + 0.6R	
	(30 to 50) g	0.14 mg + 0.6R	
	(50 to 100) g	0.29 mg + 0.6R	
	(100 to 200) g	0.58 mg + 0.6R	
	(200 to 300) g	0.87 mg + 0.6R	
	(300 to 500) g	1.4 mg + 0.6R	
	(500 to 1000) g	2.9 mg + 0.6R	
	Up to 2 lb	0.0002 lb + 0.6R	
	(2 to 10) lb	0.001 lb + 0.6R	
	(10 to 50) lb	0.0029 lb + 0.6R	
	(50 to 100) lb	0.012 lb + 0.6R	
	(100 to 500) lb	0.059 lb + 0.6R	
(500 to 1000) lb	0.11 lb + 0.6R		
(1000 to 200 000) lb	(0.11 lb/1000 lb) + 0.6R		

¹ This laboratory offers commercial calibration service and field calibration service.

- ² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.
- ³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g., resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.
- ⁴ In the statement of CMC, R represents the resolution of the unit under test.